

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1322

Name: RITTER2  
Sterilizer ID: V2313228  
07/20/2022 08:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.6 F 0.0  
2:00 155.2 F 0.0  
4:00 160.6 F 0.5  
6:00 188.7 F 4.7  
8:00 214.5 F 9.6  
10:00 228.4 F 8.3  
12:00 233.9 F 10.2  
14:00 236.0 F 8.4  
16:00 243.4 F 13.5  
18:00 242.8 F 12.4  
20:00 255.1 F 18.8  
22:00 265.9 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.6 F 29.9  
2:00 271.9 F 29.2  
3:00 272.0 F 29.8  
4:00 272.5 F 29.7

Min 271.4 F 29.1  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 212.9 F 0.0  
10:00 235.8 F 0.0  
15:00 241.7 F 0.0  
20:00 237.8 F 0.0  
25:00 224.1 F 0.0  
30:00 220.5 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 23:46  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:01:49  
CYCLE COMPLETE

07/20/2022 09:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_