
Midmark M11 - v1.0.5
Total Cycles: 3252

Name: RITTER2
Sterilizer ID: V2313228
12/22/2023 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.1 F 0.0
2:00 174.3 F 3.5
4:00 196.5 F 8.8
6:00 217.2 F 7.7
8:00 226.8 F 10.2
10:00 232.8 F 7.6
12:00 242.2 F 11.5
14:00 246.3 F 13.5
16:00 259.4 F 20.7
18:00 271.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.0 F 29.1
2:00 272.2 F 29.2
3:00 272.8 F 29.4
4:00 272.9 F 29.7

Min 271.6 F 29.1
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.2 F 0.0
5:00 234.0 F 0.0
10:00 227.0 F 0.0
15:00 193.0 F 0.0
20:00 190.2 F 0.0
25:00 188.5 F 0.0
30:00 194.3 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

12/22/2023 03:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____