
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002248
9:36 13/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:26
STERILIZING 3:22
136.0 C 220kPa 3:22
136.3 C 224kPa 3:52
136.4 C 225kPa 4:22
136.4 C 225kPa 4:52
136.3 C 223kPa 5:22
136.4 C 224kPa 5:52
136.4 C 225kPa 6:22
136.4 C 223kPa 6:52
136.4 C 223kPa 6:52
VENTING 6:52
AIR DRYING 7:43
CYCLE COMPLETE 29:49

Digital Signature #
9E73D443BC908DA6
