
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001443
8:43 09/17/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:11
PRESSURIZING 6:55
STERILIZING 8:35
136.4 C 222kPa 8:35
136.5 C 222kPa 9:05
136.3 C 222kPa 9:35
136.4 C 222kPa 10:05
136.4 C 223kPa 10:35
136.3 C 221kPa 11:05
136.0 C 219kPa 11:35
136.4 C 217kPa 12:05
VENTING 12:05
AIR DRYING 13:09
CYCLE COMPLETE 43:09

Digital Signature #
BDAA373BD4CBE526
