

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001043  
14:55 20/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:31  
PRESSURIZING 4:31  
STERILIZING 5:42  
136.0 C 221kPa 5:42  
136.4 C 225kPa 6:12  
136.5 C 224kPa 6:42  
136.2 C 223kPa 7:12  
136.3 C 224kPa 7:42  
136.4 C 225kPa 8:12  
136.1 C 223kPa 8:42  
136.2 C 222kPa 9:12  
136.2 C 222kPa 9:12  
VENTING 9:12  
AIR DRYING 10:10  
CYCLE COMPLETE 20:41  
-----

Digital Signature #  
D6CFAC821309F3B8  
-----