
Midmark M11 - v1.0.5
Total Cycles: 2660

Name: RITTER2
Sterilizer ID: V2313228
07/18/2023 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.4 F 0.0
2:00 148.5 F 0.0
4:00 160.7 F 2.9
6:00 182.9 F 7.0
8:00 205.0 F 11.9
10:00 217.6 F 6.3
12:00 231.0 F 11.3
14:00 231.7 F 8.3
16:00 241.1 F 11.4
18:00 244.6 F 13.1
20:00 257.2 F 19.9
22:00 268.5 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.6 F 29.8
2:00 271.6 F 29.0
3:00 272.4 F 30.0
4:00 272.4 F 29.6

Min 271.5 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 206.3 F 0.0
10:00 211.1 F 0.0
15:00 212.7 F 0.0
20:00 215.5 F 0.0
25:00 213.1 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:13
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

07/18/2023 10:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____