
Midmark M11 - v1.0.0
Total Cycles: 654

Name: MIDMARK
Sterilizer ID: V2178198
02/08/2021 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.2 F 0.0
2:00 143.3 F 1.1
4:00 172.8 F 4.2
6:00 188.6 F 9.5
8:00 214.7 F 5.6
10:00 226.4 F 9.1
12:00 234.5 F 9.2
14:00 241.8 F 12.3
16:00 241.5 F 10.7
18:00 255.4 F 17.7
20:00 267.4 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.4 F 28.9
2:00 272.7 F 28.8
3:00 272.5 F 28.5
4:00 272.2 F 28.4

Min 272.0 F 28.4
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 217.9 F 0.0
10:00 206.4 F 0.0
15:00 229.2 F 0.0
20:00 236.6 F 0.0
25:00 229.1 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 21:54
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 00:59:45
CYCLE COMPLETE

02/08/2021 02:03 PM
ACCEPTED: X
REJECTED:
Operator: U01