

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2339

Name: RITTER2  
Sterilizer ID: V2313228  
05/04/2023 11:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
2:00 198.8 F 4.2  
4:00 206.8 F 9.7  
6:00 221.4 F 6.8  
8:00 235.1 F 11.8  
10:00 236.5 F 10.0  
12:00 235.8 F 8.7  
14:00 250.8 F 16.2  
16:00 263.0 F 23.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.1  
1:00 272.8 F 29.8  
2:00 271.5 F 29.1  
3:00 272.4 F 29.6  
4:00 271.4 F 29.1  
  
Min 271.4 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
5:00 221.3 F 0.0  
10:00 225.5 F 0.0  
15:00 229.6 F 0.0  
20:00 224.4 F 0.0  
25:00 225.5 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 2:17  
HEATING: 17:60  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:57:33  
CYCLE COMPLETE

05/04/2023 12:24 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_