
Midmark M11 - v1.0.5
Total Cycles: 4426

Name: RITTER2
Sterilizer ID: V2313228
03/13/2025 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.1 F 0.0
2:00 106.5 F 0.8
4:00 115.3 F 2.7
6:00 132.1 F 5.6
8:00 176.4 F 8.7
10:00 181.9 F 12.4
12:00 205.7 F 5.2
14:00 216.5 F 9.0
16:00 227.9 F 13.1
18:00 230.9 F 9.7
20:00 235.1 F 9.9
22:00 242.7 F 13.1
24:00 253.7 F 18.9
26:00 263.7 F 24.9
28:00 272.7 F 30.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 30.2
1:00 271.6 F 30.6
2:00 272.8 F 30.8
3:00 271.7 F 30.3
4:00 271.3 F 30.3

Min 271.3 F 30.1
Max 273.3 F 31.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 196.3 F 0.0
10:00 200.2 F 0.0
15:00 203.8 F 0.0
20:00 202.6 F 0.0
25:00 209.5 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 28:10
STERILIZING: 4:00
VENTING: 4:01
DRYING: 30:00
TOTAL CYCLE: 01:07:12
CYCLE COMPLETE

03/13/2025 04:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____