

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002660  
14:06 07/04/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:26  
STERILIZING 3:34  
136.4 C 227kPa 3:34  
136.4 C 223kPa 4:04  
136.1 C 224kPa 4:34  
136.4 C 226kPa 5:04  
136.4 C 226kPa 5:34  
136.4 C 226kPa 6:04  
136.4 C 225kPa 6:34  
136.4 C 225kPa 7:04  
136.4 C 225kPa 7:04  
VENTING 7:04  
AIR DRYING 7:58  
CYCLE COMPLETE 37:58  
-----

Digital Signature #  
B35C5AA521A796DA  
-----