
Midmark M11 - v1.0.0
Total Cycles: 748

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2021 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.2 F 0.0
2:00 129.1 F 1.1
4:00 152.5 F 3.6
6:00 182.3 F 7.4
8:00 204.1 F 13.2
10:00 216.5 F 6.7
12:00 230.1 F 11.9
14:00 232.9 F 8.6
16:00 239.2 F 10.2
18:00 245.1 F 13.1
20:00 257.5 F 19.9
22:00 269.2 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.5 F 29.6
2:00 272.1 F 28.9
3:00 271.5 F 28.8
4:00 272.5 F 29.5

Min 271.2 F 28.4
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 211.4 F 0.0
10:00 210.0 F 0.0
15:00 209.1 F 0.0
20:00 225.0 F 0.0
25:00 229.5 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:06
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:01:11
CYCLE COMPLETE

03/25/2021 01:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____