

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2583

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2022 10:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.9 F 0.0  
2:00 157.2 F 1.4  
4:00 166.4 F 5.2  
6:00 190.1 F 9.4  
8:00 218.8 F 7.6  
10:00 218.2 F 7.3  
12:00 232.5 F 11.9  
14:00 234.8 F 8.4  
16:00 241.4 F 11.3  
18:00 244.9 F 12.4  
20:00 257.2 F 18.9  
22:00 268.2 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 271.5 F 28.4  
2:00 271.7 F 28.6  
3:00 272.5 F 28.9  
4:00 272.1 F 29.0

Min 271.2 F 28.4  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.7 F 0.0  
5:00 231.0 F 0.0  
10:00 204.4 F 0.0  
15:00 202.6 F 0.0  
20:00 230.5 F 0.0  
25:00 234.1 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 23:59  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:02:06  
CYCLE COMPLETE

11/21/2022 11:20 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_