
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001557
10:15 11/04/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:11
PRESSURIZING 7:04
STERILIZING 8:44
136.0 C 220kPa 8:44
136.1 C 219kPa 9:14
136.2 C 219kPa 9:44
136.4 C 222kPa 10:15
136.4 C 220kPa 10:45
136.4 C 223kPa 11:15
136.2 C 221kPa 11:45
136.4 C 221kPa 12:14
VENTING 12:14
AIR DRYING 13:20
CYCLE COMPLETE 43:20

Digital Signature #
104BBD37F50C866A
