
Midmark M11 - v1.0.0
Total Cycles: 3726

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2023 12:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.4 F 0.0
2:00 161.5 F 0.9
4:00 186.8 F 3.9
6:00 207.3 F 8.3
8:00 227.6 F 12.2
10:00 232.9 F 8.8
12:00 241.3 F 11.0
14:00 242.5 F 11.9
16:00 240.5 F 10.5
18:00 254.1 F 17.3
20:00 265.5 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.1 F 28.9
2:00 271.9 F 28.4
3:00 271.5 F 28.6
4:00 272.9 F 29.2

Min 270.9 F 28.0
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.6 F 0.0
5:00 228.5 F 0.0
10:00 216.5 F 0.0
15:00 198.9 F 0.0
20:00 204.2 F 0.0
25:00 204.7 F 0.0
30:00 203.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

09/06/2023 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____