
Midmark M11 - v1.0.5
Total Cycles: 2700

Name: RITTER2
Sterilizer ID: V2313228
07/25/2023 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.0 F 0.0
2:00 183.1 F 2.4
4:00 194.8 F 7.1
6:00 211.4 F 11.7
8:00 223.5 F 6.6
10:00 236.9 F 11.4
12:00 234.7 F 9.0
14:00 239.9 F 10.5
16:00 247.6 F 14.2
18:00 259.1 F 20.6
20:00 269.6 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.3
2:00 271.3 F 28.8
3:00 272.6 F 29.4
4:00 271.5 F 28.7

Min 271.3 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 212.9 F 0.0
10:00 224.2 F 0.0
15:00 223.8 F 0.0
20:00 219.8 F 0.0
25:00 217.5 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

07/25/2023 05:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____