
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001675
13:07 01/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:32
STERILIZING 3:39
136.4 C 222kPa 3:39
136.4 C 225kPa 4:09
136.4 C 224kPa 4:39
136.4 C 220kPa 5:09
136.4 C 225kPa 5:39
136.4 C 225kPa 6:09
136.4 C 224kPa 6:39
CYCLE ABORTED 7:00
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:00

Digital Signature #
D21DA7FCCA892ED3
