
Midmark M11 - v1.0.5
Total Cycles: 2489

Name: RITTER2
Sterilizer ID: V2313228
06/08/2023 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 192.9 F 4.8
4:00 211.1 F 10.8
6:00 226.6 F 6.7
8:00 238.3 F 12.8
10:00 240.4 F 11.2
12:00 239.5 F 10.7
14:00 254.5 F 18.0
16:00 266.8 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.3 F 29.2
2:00 271.5 F 29.1
3:00 272.2 F 29.2
4:00 271.3 F 28.9

Min 271.2 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 226.7 F 0.0
10:00 228.6 F 0.0
15:00 231.7 F 0.0
20:00 224.5 F 0.0
25:00 229.5 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 17:26
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:56:56
CYCLE COMPLETE

06/08/2023 03:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____