

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.8uS / 0.5ppm  
CYCLE NUMBER 003758  
8:23 06/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:11  
PRESSURIZING 3:13  
STERILIZING 4:26  
136.0 C 220kPa 4:26  
136.2 C 223kPa 4:56  
136.4 C 223kPa 5:26  
136.4 C 223kPa 5:56  
136.4 C 221kPa 6:26  
136.4 C 223kPa 6:56  
136.4 C 222kPa 7:26  
136.4 C 222kPa 7:56  
136.4 C 222kPa 7:56  
VENTING 7:56  
AIR DRYING 8:44  
CYCLE COMPLETE 9:10

-----  
Digital Signature #  
A9B545B515DC2FB3  
-----