
Midmark M11 - v1.0.0
Total Cycles: 936

Name: RITTER 1
Sterilizer ID: V2178198
06/24/2021 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.7 F 0.0
2:00 188.3 F 2.9
4:00 194.6 F 6.7
6:00 212.0 F 11.3
8:00 219.9 F 5.1
10:00 232.7 F 9.9
12:00 235.7 F 9.2
14:00 241.6 F 12.0
16:00 239.9 F 10.3
18:00 252.0 F 16.3
20:00 262.6 F 22.4
22:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.1
2:00 272.1 F 28.6
3:00 271.5 F 28.4
4:00 272.5 F 29.1

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 209.4 F 0.0
10:00 198.2 F 0.0
15:00 197.6 F 0.0
20:00 195.3 F 0.0
25:00 194.1 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:21
CYCLE COMPLETE

06/24/2021 12:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____