

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2855

Name: RITTER 1  
Sterilizer ID: V2178198  
02/03/2023 10:54 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.0 F 0.0  
2:00 175.5 F 2.6  
4:00 183.6 F 7.1  
6:00 206.6 F 13.1  
8:00 225.8 F 7.7  
10:00 237.1 F 10.4  
12:00 243.3 F 12.4  
14:00 242.5 F 11.3  
16:00 257.5 F 19.0  
18:00 270.1 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.9  
1:00 271.8 F 27.8  
2:00 272.4 F 28.3  
3:00 272.9 F 28.6  
  
Min 271.4 F 27.8  
Max 272.9 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 226.4 F 0.0  
5:00 224.3 F 0.0  
Dry Cycle Aborted

FILLING: 1:42  
HEATING: 18:46  
STERILIZING: 3:00  
VENTING: 2:55  
DRYING: 8:33  
TOTAL CYCLE: 00:34:57

02/03/2023 11:28 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_