
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008353
14:42 25/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:16
STERILIZING 4:14
136.4 C 224kPa 4:14
136.3 C 224kPa 4:45
136.0 C 226kPa 5:15
136.4 C 225kPa 5:45
136.4 C 225kPa 6:15
136.4 C 223kPa 6:45
136.5 C 224kPa 7:15
136.4 C 225kPa 7:45
136.4 C 225kPa 7:45
VENTING 7:45
AIR DRYING 8:38
CYCLE COMPLETE 14:02

Digital Signature #
A5D6BF3C5336188F
