

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1037

Name: RITTER 1  
Sterilizer ID: V2178198  
07/21/2021 10:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 219.4 F 0.0  
2:00 215.7 F 4.1  
4:00 220.3 F 9.4  
6:00 229.6 F 6.9  
8:00 235.3 F 10.5  
10:00 237.3 F 7.5  
12:00 242.2 F 11.5  
14:00 245.9 F 13.1  
16:00 258.1 F 19.7  
18:00 268.9 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.3 F 28.9  
2:00 271.4 F 28.3  
3:00 272.4 F 29.0  
4:00 271.8 F 28.3  
  
Min 271.4 F 28.2  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.8 F 0.0  
5:00 232.2 F 0.0  
10:00 228.6 F 0.0  
15:00 229.3 F 0.0  
20:00 228.9 F 0.0  
25:00 229.4 F 0.0  
30:00 213.7 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 19:07  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:20  
CYCLE COMPLETE

07/21/2021 11:30 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01