
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001006
10:33 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:11
STERILIZING 3:09
136.0 C 229kPa 3:09
136.1 C 232kPa 3:39
136.3 C 232kPa 4:09
136.1 C 230kPa 4:39
136.4 C 233kPa 5:09
136.1 C 231kPa 5:39
136.4 C 231kPa 6:09
136.4 C 233kPa 6:39
136.4 C 233kPa 6:39
VENTING 6:39
AIR DRYING 7:30
CYCLE COMPLETE 7:34

Digital Signature #
0DE2F8CE8F9C719A
