
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001410
7:45 09/09/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 4:37
STERILIZING 6:07
136.0 C 220kPa 6:07
136.4 C 222kPa 6:37
136.4 C 223kPa 7:07
136.0 C 220kPa 7:37
136.1 C 220kPa 8:08
136.3 C 220kPa 8:38
136.4 C 222kPa 9:08
135.9 C 220kPa 9:37
VENTING 9:38
AIR DRYING 10:44
CYCLE COMPLETE 11:00

Digital Signature #
F470F46EB5A16932
