
Midmark M11 - v1.0.0
Total Cycles: 4771

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2024 02:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.4 F 0.0
2:00 121.4 F 1.6
4:00 149.4 F 4.4
6:00 168.8 F 7.6
8:00 207.6 F 11.7
10:00 213.3 F 3.7
12:00 222.7 F 7.7
14:00 232.7 F 11.8
16:00 231.2 F 7.5
18:00 243.2 F 12.7
20:00 240.5 F 10.6
22:00 252.0 F 16.4
24:00 262.6 F 22.4
26:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 271.9 F 28.8
2:00 272.4 F 28.8
3:00 271.3 F 28.2
4:00 271.8 F 28.8

Min 271.1 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 245.1 F 0.0
10:00 229.1 F 0.0
15:00 229.0 F 0.0
20:00 223.0 F 0.0
25:00 215.5 F 0.0
30:00 207.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:21
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:05:06
CYCLE COMPLETE

07/26/2024 03:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____