
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 003771
8:00 07/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:03
PRESSURIZING 4:04
STERILIZING 5:01
136.4 C 223kPa 5:01
135.9 C 219kPa 5:31
136.4 C 223kPa 6:01
136.3 C 222kPa 6:31
136.2 C 222kPa 7:01
136.5 C 224kPa 7:31
136.4 C 223kPa 8:01
136.4 C 220kPa 8:31
136.4 C 220kPa 8:31
VENTING 8:31
AIR DRYING 9:23
CYCLE COMPLETE 13:10

Digital Signature #
A9BF4A2B3DF5BFC6
