
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004927
11:25 09/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:46
PRESSURIZING 1:49
STERILIZING 2:55
136.4 C 221kPa 2:55
136.4 C 225kPa 3:25
136.4 C 222kPa 3:55
136.2 C 223kPa 4:25
136.4 C 222kPa 4:55
136.3 C 217kPa 5:25
136.4 C 223kPa 5:55
136.4 C 223kPa 6:25
136.4 C 223kPa 6:25
VENTING 6:25
AIR DRYING 7:16
CYCLE COMPLETE 9:36

Digital Signature #
524BD68C525AADA6
