
Midmark M11 - v1.0.5
Total Cycles: 1640

Name: RITTER2
Sterilizer ID: V2313228
10/21/2022 03:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.5 F 0.0
2:00 201.2 F 4.0
4:00 208.2 F 9.1
6:00 221.6 F 8.0
8:00 230.1 F 10.1
10:00 233.1 F 8.2
12:00 242.3 F 12.1
14:00 245.3 F 13.4
16:00 258.1 F 20.3
18:00 269.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.4 F 29.7
2:00 271.6 F 28.9
3:00 271.9 F 29.8
4:00 272.2 F 29.5

Min 271.1 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 230.9 F 0.0
10:00 232.1 F 0.0
15:00 230.6 F 0.0
20:00 227.6 F 0.0
25:00 230.2 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

10/21/2022 04:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____