
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006341
9:43 25/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:04
STERILIZING 3:09
136.5 C 223kPa 3:09
136.2 C 222kPa 3:39
136.4 C 223kPa 4:09
136.1 C 217kPa 4:39
136.0 C 219kPa 5:09
136.4 C 220kPa 5:39
135.9 C 219kPa 6:09
136.1 C 221kPa 6:39
136.1 C 221kPa 6:39
VENTING 6:39
AIR DRYING 7:30
CYCLE COMPLETE 8:24

Digital Signature #
B036ADF2C7AD53BE
