
Midmark M11 - v1.0.5
Total Cycles: 2507

Name: RITTER2
Sterilizer ID: V2313228
06/13/2023 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.7 F 0.0
2:00 192.0 F 3.9
4:00 213.6 F 9.4
6:00 222.6 F 7.8
8:00 234.7 F 11.0
10:00 234.1 F 8.8
12:00 238.9 F 9.8
14:00 248.2 F 14.9
16:00 260.8 F 22.0
18:00 272.1 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.5
2:00 272.0 F 29.7
3:00 272.2 F 29.4
4:00 271.9 F 29.6

Min 271.4 F 29.0
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 218.2 F 0.0
10:00 236.4 F 0.0
15:00 229.8 F 0.0
20:00 227.7 F 0.0
25:00 225.3 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 18:29
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

06/13/2023 11:34 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____