
Midmark M11 - v1.0.0
Total Cycles: 2181

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2022 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.2 F 0.0
2:00 195.8 F 3.3
4:00 200.5 F 8.7
6:00 221.2 F 6.8
8:00 231.3 F 9.6
10:00 235.3 F 8.7
12:00 244.7 F 13.0
14:00 242.9 F 11.7
16:00 255.8 F 18.5
18:00 267.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.3 F 28.9
2:00 272.1 F 28.9
3:00 272.3 F 28.7
4:00 272.4 F 29.1

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 225.5 F 0.0
10:00 225.9 F 0.0
15:00 244.0 F 0.0
20:00 232.2 F 0.0
25:00 226.7 F 0.0
30:00 236.9 F 0.0
Drying Complete

FILLING: 2:27
HEATING: 19:26
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

07/26/2022 03:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____