
Midmark M11 - v1.0.0
Total Cycles: 4414

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2024 04:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.3 F 0.0
2:00 148.9 F 2.8
4:00 171.8 F 6.4
6:00 193.7 F 10.9
8:00 219.0 F 4.1
10:00 222.9 F 9.5
12:00 235.1 F 9.0
14:00 240.6 F 11.2
16:00 237.2 F 9.2
18:00 251.1 F 15.9
20:00 263.0 F 22.8
22:00 271.9 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.1 F 28.7
2:00 272.7 F 29.1
3:00 272.1 F 29.1
4:00 272.0 F 28.7

Min 271.4 F 28.5
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 228.6 F 0.0
10:00 223.7 F 0.0
15:00 217.8 F 0.0
20:00 230.9 F 0.0
25:00 231.0 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

03/27/2024 05:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____