
Midmark M11 - v1.0.0
Total Cycles: 2556

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2022 05:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.1 F 0.0
2:00 197.4 F 4.8
4:00 221.8 F 10.1
6:00 227.2 F 5.6
8:00 238.1 F 11.5
10:00 236.9 F 9.1
12:00 240.5 F 9.3
14:00 249.3 F 14.7
16:00 261.8 F 21.7
18:00 272.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.7
2:00 271.4 F 28.0
3:00 272.6 F 28.8
4:00 271.8 F 28.2

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 234.7 F 0.0
10:00 224.6 F 0.0
15:00 226.3 F 0.0
20:00 227.2 F 0.0
25:00 230.7 F 0.0
30:00 229.6 F 0.0

Drying Complete

FILLING: 2:33
HEATING: 18:21
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:06
CYCLE COMPLETE

11/14/2022 06:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____