

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1237

Name: RITTER 1  
Sterilizer ID: V2178198  
10/06/2021 02:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.4 F 0.0  
2:00 170.9 F 1.6  
4:00 200.3 F 4.6  
6:00 215.5 F 8.9  
8:00 225.5 F 10.3  
10:00 227.5 F 7.2  
12:00 237.9 F 11.9  
14:00 234.0 F 8.5  
16:00 242.4 F 11.3  
18:00 245.0 F 12.7  
20:00 256.9 F 19.0  
22:00 267.6 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.7 F 29.0  
2:00 272.3 F 28.5  
3:00 271.4 F 28.1  
4:00 272.4 F 29.0

Min 271.4 F 28.1  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.9 F 0.0  
5:00 222.9 F 0.0  
10:00 222.5 F 0.0  
15:00 221.2 F 0.0  
20:00 214.8 F 0.0  
25:00 221.0 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 23:28  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:01:53  
CYCLE COMPLETE

10/06/2021 03:51 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01