
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005737
10:29 12/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:33
STERILIZING 3:36
136.5 C 225kPa 3:36
136.2 C 222kPa 4:07
136.4 C 225kPa 4:37
136.2 C 222kPa 5:07
135.9 C 221kPa 5:37
136.3 C 224kPa 6:07
136.1 C 222kPa 6:37
136.1 C 222kPa 7:07
136.1 C 222kPa 7:07
VENTING 7:07
AIR DRYING 7:54
CYCLE COMPLETE 8:20

Digital Signature #
4C37C7174D14D443
