
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000770
13:20 28/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:10
STERILIZING 4:21
136.4 C 233kPa 4:21
136.4 C 233kPa 4:51
136.4 C 233kPa 5:21
136.3 C 232kPa 5:51
136.1 C 231kPa 6:21
136.4 C 233kPa 6:51
136.5 C 233kPa 7:21
136.3 C 231kPa 7:51
136.3 C 231kPa 7:51
VENTING 7:51
AIR DRYING 8:46
CYCLE COMPLETE 8:48

Digital Signature #
28E3493120265F2D
