
Midmark M11 - v1.0.5
Total Cycles: 3281

Name: RITTER2
Sterilizer ID: V2313228
01/10/2024 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.2 F 0.0
2:00 168.2 F 1.5
4:00 181.6 F 5.4
6:00 201.4 F 9.7
8:00 216.7 F 8.5
10:00 224.9 F 9.1
12:00 234.3 F 10.9
14:00 240.8 F 12.8
16:00 241.9 F 12.2
18:00 254.8 F 18.9
20:00 266.3 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.6 F 30.1
2:00 271.6 F 29.4
3:00 271.8 F 30.0
4:00 272.4 F 30.0

Min 271.1 F 29.3
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 208.8 F 0.0
10:00 207.4 F 0.0
15:00 200.7 F 0.0
20:00 198.2 F 0.0
25:00 197.1 F 0.0
30:00 200.2 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:40
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:00:51
CYCLE COMPLETE

01/10/2024 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____