
Midmark M11 - v1.0.0
Total Cycles: 3655

Name: RITTER 1
Sterilizer ID: V2178198
08/15/2023 12:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.2 F 0.0
2:00 190.4 F 4.3
4:00 210.4 F 10.1
6:00 224.7 F 5.9
8:00 239.4 F 11.3
10:00 235.6 F 8.9
12:00 239.2 F 9.7
14:00 248.6 F 14.4
16:00 261.1 F 21.4
18:00 272.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 271.6 F 28.1
2:00 271.8 F 28.1
3:00 271.9 F 28.3
4:00 272.3 F 28.5

Min 271.5 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 240.9 F 0.0
10:00 226.8 F 0.0
15:00 224.0 F 0.0
20:00 218.5 F 0.0
25:00 232.3 F 0.0
30:00 227.7 F 0.0

Drying Complete

FILLING: 2:17
HEATING: 18:23
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

08/15/2023 01:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____