

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076

Unit # : 000

WATER QUALITY

1.9uS / 1.2ppm

CYCLE NUMBER 000028

9:46 10/08/2020

RUBBER AND PLASTICS

121 C for 15 min.

CYCLE START 0:00

CONDITIONING 2:57

PRESSURIZING 8:01

STERILIZING 8:46

123.4 C 120kPa 8:46

123.3 C 119kPa 9:16

123.4 C 120kPa 9:46

123.3 C 120kPa 10:16

123.5 C 120kPa 10:46

123.4 C 120kPa 11:16

123.4 C 120kPa 11:46

123.3 C 119kPa 12:16

122.8 C 120kPa 12:46

123.2 C 119kPa 13:16

123.4 C 120kPa 13:46

123.4 C 120kPa 14:16

123.5 C 120kPa 14:46

123.3 C 120kPa 15:16

123.5 C 120kPa 15:46

123.4 C 120kPa 16:16

123.4 C 120kPa 16:46

123.4 C 120kPa 17:16

123.4 C 120kPa 17:46

123.2 C 119kPa 18:16

123.4 C 120kPa 18:46

123.4 C 120kPa 19:17

123.4 C 120kPa 19:47

123.3 C 120kPa 20:17

123.4 C 120kPa 20:47

123.4 C 120kPa 21:17

123.3 C 120kPa 21:47

123.2 C 119kPa 22:17

123.5 C 118kPa 22:47

123.4 C 120kPa 23:17

123.3 C 120kPa 23:46

VENTING 23:46

AIR DRYING 24:31  
CYCLE COMPLETE 54:31

-----

Digital Signature #  
5A52F1A9255C63C6

-----