
Midmark M11 - v1.0.0
Total Cycles: 1836

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2022 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 176.7 F 2.9
4:00 185.2 F 7.4
6:00 205.4 F 12.5
8:00 224.1 F 6.1
10:00 236.4 F 11.1
12:00 233.3 F 7.6
14:00 245.1 F 12.3
16:00 245.1 F 12.5
18:00 257.8 F 19.2
20:00 269.2 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.2 F 28.3
2:00 272.5 F 28.6
3:00 272.6 F 28.9
4:00 271.7 F 28.4

Min 271.7 F 28.2
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 236.6 F 0.0
10:00 198.6 F 0.0
15:00 194.9 F 0.0
20:00 194.7 F 0.0
25:00 195.3 F 0.0
30:00 211.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 21:41
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

04/18/2022 03:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____