
Midmark M11 - v1.0.5
Total Cycles: 3966

Name: RITTER2
Sterilizer ID: V2313228
09/23/2024 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 71.7 F 0.0
4:00 78.6 F 0.0
6:00 95.3 F 0.1
8:00 144.9 F 0.3
10:00 176.1 F 0.5
12:00 198.2 F 1.8
14:00 204.7 F 5.0
16:00 215.8 F 8.1
18:00 224.5 F 11.5
20:00 227.6 F 6.7
22:00 237.6 F 11.8
24:00 236.4 F 9.5
26:00 239.1 F 9.9
28:00 247.9 F 14.4
30:00 259.2 F 20.6
32:00 269.4 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 29.3
2:00 271.5 F 28.7
3:00 272.5 F 29.4
4:00 271.8 F 28.7

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 185.6 F 0.0
10:00 188.5 F 0.0
15:00 183.5 F 0.0
20:00 180.9 F 0.0
25:00 179.6 F 0.0
30:00 182.3 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:05
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:11:27
CYCLE COMPLETE

09/23/2024 10:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____