
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000270
9:17 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:13
STERILIZING 4:50
136.3 C 232kPa 4:50
136.4 C 233kPa 5:20
136.3 C 232kPa 5:50
136.4 C 233kPa 6:20
136.3 C 232kPa 6:50
136.4 C 233kPa 7:20
136.3 C 232kPa 7:50
136.4 C 233kPa 8:20
136.4 C 233kPa 8:20
VENTING 8:20
AIR DRYING 9:27
CYCLE COMPLETE 24:27

Digital Signature #
CA757D6A0E9602CA
