
Midmark M11 - v1.0.0
Total Cycles: 4002

Name: RITTER 1
Sterilizer ID: V2178198
11/09/2023 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.0 F 0.0
2:00 159.1 F 3.1
4:00 182.8 F 7.2
6:00 203.1 F 11.9
8:00 225.8 F 6.2
10:00 231.9 F 11.5
12:00 233.7 F 8.4
14:00 241.0 F 10.9
16:00 245.7 F 13.1
18:00 257.9 F 19.6
20:00 268.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.1
2:00 271.6 F 28.9
3:00 272.4 F 29.1
4:00 271.3 F 28.7

Min 271.3 F 28.6
Max 273.2 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.3 F 0.0
5:00 209.9 F 0.0
10:00 198.2 F 0.0
15:00 197.4 F 0.0
20:00 195.2 F 0.0
25:00 193.0 F 0.0
30:00 190.9 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:55
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:47
CYCLE COMPLETE

11/09/2023 02:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____