

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3547

Name: RITTER2  
Sterilizer ID: V2313228  
04/22/2024 04:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
2:00 180.7 F 2.6  
4:00 190.7 F 7.2  
6:00 206.8 F 12.9  
8:00 226.1 F 7.7  
10:00 238.8 F 13.4  
12:00 239.7 F 11.2  
14:00 238.6 F 10.1  
16:00 252.6 F 17.1  
18:00 264.5 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 272.3 F 29.1  
2:00 272.6 F 29.6  
3:00 271.6 F 29.0  
4:00 272.3 F 29.2  
  
Min 271.6 F 28.8  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.7 F 0.0  
5:00 229.9 F 0.0  
10:00 228.9 F 0.0  
15:00 226.4 F 0.0  
20:00 224.3 F 0.0  
25:00 223.2 F 0.0  
30:00 223.3 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 19:50  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:44  
CYCLE COMPLETE

04/22/2024 05:13 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_