
Midmark M11 - v1.0.5
Total Cycles: 4159

Name: RITTER2
Sterilizer ID: V2313228
11/26/2024 04:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.7 F 0.0
2:00 115.0 F 0.0
4:00 112.4 F 0.1
6:00 155.9 F 0.4
8:00 188.2 F 2.3
10:00 205.2 F 6.5
12:00 219.1 F 10.9
14:00 225.0 F 6.2
16:00 237.0 F 11.5
18:00 235.5 F 9.2
20:00 238.5 F 9.6
22:00 249.0 F 15.1
24:00 261.6 F 22.3
26:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.8 F 29.4
2:00 272.3 F 29.4
3:00 271.8 F 28.7
4:00 272.7 F 29.4

Min 271.6 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.1 F 0.0
5:00 207.6 F 0.0
10:00 216.9 F 0.0
15:00 215.6 F 0.0
20:00 211.9 F 0.0
25:00 219.2 F 0.0
30:00 224.4 F 0.0

Drying Complete

FILLING: 0:57
HEATING: 26:09
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:04:22
CYCLE COMPLETE

11/26/2024 05:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____