
Midmark M11 - v1.0.5
Total Cycles: 944

Name: RITTER2
Sterilizer ID: V2313228
04/01/2022 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.1 F 0.0
2:00 192.7 F 3.1
4:00 209.6 F 8.8
6:00 230.1 F 9.1
8:00 235.1 F 11.9
10:00 238.5 F 10.6
12:00 238.4 F 10.2
14:00 253.4 F 17.5
16:00 265.6 F 25.0
18:00 273.0 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.6
1:00 272.2 F 29.5
2:00 272.9 F 29.8
3:00 272.7 F 30.2
4:00 271.7 F 29.6

Min 271.7 F 29.4
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 223.7 F 0.0
10:00 206.5 F 0.0
15:00 213.4 F 0.0
20:00 193.4 F 0.0
25:00 190.4 F 0.0
30:00 191.4 F 0.0
Drying Complete

FILLING: 2:56
HEATING: 18:17
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

04/01/2022 11:56 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____