
Midmark M11 - v1.0.0
Total Cycles: 2788

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2023 04:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.8 F 0.0
2:00 175.1 F 3.2
4:00 202.9 F 7.2
6:00 214.2 F 13.0
8:00 223.2 F 7.2
10:00 235.5 F 12.5
12:00 237.2 F 9.8
14:00 235.9 F 8.6
16:00 249.7 F 15.3
18:00 262.4 F 22.4
20:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.3
2:00 272.3 F 29.0
3:00 271.7 F 28.5
4:00 271.2 F 28.4

Min 271.2 F 28.3
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.3 F 0.0
5:00 234.2 F 0.0
10:00 229.1 F 0.0
15:00 229.3 F 0.0
20:00 227.1 F 0.0
25:00 229.1 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:01
CYCLE COMPLETE

01/17/2023 05:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____