
Midmark M11 - v1.0.0
Total Cycles: 606

Name: MIDMARK
Sterilizer ID: V2178198
01/14/2021 04:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.9 F 0.0
2:00 120.8 F 0.1
4:00 135.3 F 1.9
6:00 173.8 F 5.2
8:00 225.4 F 9.7
10:00 229.4 F 7.4
12:00 228.5 F 8.8
14:00 239.6 F 11.3
16:00 240.8 F 11.1
18:00 237.3 F 9.2
20:00 251.3 F 16.0
22:00 263.4 F 22.9
24:00 270.7 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.5 F 29.1
2:00 272.2 F 28.7
3:00 271.7 F 28.9
4:00 272.3 F 29.2

Min 271.7 F 28.7
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 242.1 F 0.0
10:00 230.3 F 0.0
15:00 229.9 F 0.0
20:00 225.4 F 0.0
25:00 229.9 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:35
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:51
CYCLE COMPLETE

01/14/2021 05:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____