

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000190  
14:56 06/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:13  
PRESSURIZING 4:13  
STERILIZING 5:30  
136.4 C 224kPa 5:30  
136.0 C 220kPa 6:00  
136.3 C 224kPa 6:30  
136.3 C 222kPa 7:00  
136.0 C 222kPa 7:30  
136.3 C 220kPa 8:00  
136.4 C 224kPa 8:30  
136.4 C 224kPa 9:00  
136.4 C 224kPa 9:00  
VENTING 9:00  
AIR DRYING 10:01  
CYCLE COMPLETE 10:03

-----  
Digital Signature #  
F729B29173E74F45  
-----