
Midmark M11 - v1.0.5
Total Cycles: 1216

Name: RITTER2
Sterilizer ID: V2313228
06/24/2022 07:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.6 F 0.0
2:00 75.7 F 0.0
4:00 107.9 F 0.0
6:00 141.2 F 0.0
8:00 172.4 F 0.2
10:00 173.2 F 2.5
12:00 209.2 F 6.0
14:00 220.3 F 9.6
16:00 231.4 F 13.3
18:00 225.3 F 8.0
20:00 234.6 F 12.4
22:00 236.1 F 9.4
24:00 240.0 F 10.1
26:00 246.6 F 13.6
28:00 257.7 F 19.7
30:00 267.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 271.4 F 29.1
2:00 272.4 F 29.2
3:00 271.3 F 28.6
4:00 271.5 F 29.3

Min 271.0 F 28.5
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 190.7 F 0.0
10:00 207.6 F 0.0
15:00 214.5 F 0.0
20:00 216.0 F 0.0
25:00 215.5 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:17
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:14
CYCLE COMPLETE

06/24/2022 08:14 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____